

ABC-108 Freezer Farm Guidelines

- New (or less than 1-3 years old) -80's ONLY
- No Cylinder supplemental backup allowed
- Long Term Storage ONLY: this is not meant for supplemental or backup storage
- Access should be limited to 12 or less visits/year
- Freezers should be on a yearly Preventative Maintenance Service Contract
- Yearly regulatory audits sent to Inst Dir for approval of continued use
- Access to be requested and approved through Chris Incarvito or Stephen Luzietti
- If temperature monitoring (Smartvue) is needed, the institute is responsible